

LIFEPAK® 1000 defibrillator power options

Data sheet

The rugged LIFEPAK 1000 defibrillator is an easy-to-use automated external defibrillator (AED) from the leader in defibrillation technology. But it's also a defibrillator powerful and adaptable enough for professional responders, featuring advanced capabilities that can help improve lifesaving outcomes and speed the transition of cardiac patients to the next level of critical care.



LIFEPAK 1000 rechargeable battery



LIFEPAK 1000
non-rechargeable battery

Power Options for the LIFEPAK 1000

The LIFEPAK 1000 defibrillator offers options that provide you with the flexibility to choose how you would like to manage power requirements for your device. You can choose between a rechargeable and non-rechargeable battery. With the Lithium-ion (Li-ion) rechargeable or Lithium Manganese Dioxide (Li MnO₂) non-rechargeable battery for the 1000, you will have the power you need.

Which Battery is Right for You?

Rechargeable battery is ideal for:

- Emergency response systems with frequent use of both defibrillation and 3-lead ECG monitoring.
- Emergency response systems also using the LIFEPAK 1000 for training simulation.
- Batteries are rechargeable within 4.5 hours for multiple uses.

Non-rechargeable battery is ideal for:

- Emergency response systems with infrequent device use due to lower call volumes.
- Public access users with infrequent device use.

LIFEPAK 1000 defibrillator power options provide:

Simplicity

- Simple to maintain—no conditioning or calibration required.
- Battery gauge displays the state of charge on each battery at the touch of a button to help you ensure you always have the power you need in an emergency.

Confidence in Our Products

- The LIFEPAK 1000 AED with battery installed has an IP55 rating for protection against dusty and wet environments.
- Stryker batteries are tested, verified, and certified to work with LIFEPAK devices.
- Batteries offered by third party vendors are not tested, verified or certified by Stryker for use with LIFEPAK devices.

Specifications

Non-rechargeable Batteries

Type: Lithium Manganese Dioxide (Li/MnO₂), 12.0 V, 4.5 Ah

Capacity: Typically will provide 440 200-joule discharges or 1030 minutes of operating time with a new battery (370 200-joule shocks or 900 minutes of operating time at 0°C (32°F)).

Weight: 0.45 kg (1.0 lb)

Shelf Life: (prior to installation) After the battery is stored for 5 years at 20° to 30°C, the device will provide 48 months of standby life.

Standby Life: A new battery provides device power for 5 years.

Low Battery Indicator: At least 30 200-joule shocks or 75 minutes of operating time remain when low battery is first indicated.

Rechargeable Batteries

Type: Lithium-ion, 11.1 V, 4.8 Ah, 5.3 Wh

Capacity: Typically will provide 261 200-joule discharges or 608 minutes of operating time with a new fully-charged battery (247 200-joule shocks or 576 minutes of operating time at 0°C (32°F)).

Battery Charging Time: Within 4.5 hours

Weight: 0.45 kg (1.0 lb), maximum

Standby Life: A new fully-charged battery provides device power for 6 months.

Low Battery Indicator: At least 30 200-joule shocks or 75 minutes of operating time remain when low battery is first indicated.

Battery Charger

Supported Battery: Lithium-ion Rechargeable Battery, 11.1 V, 4.8 Ah, 53 Wh

Electrical: External Power Supply: 100-240VAC, 50/60Hz

Temperature: Operating: 0°C to 40°C

Storage: -30°C to 70°C

Charge Time: Within 4.5 hours

Charge: Constant Current/Constant Voltage within temperature limits

Length: 270 mm

Width: 97 mm

Height: 92 mm

Weight: 0.5 kg

LIFEPAK 1000 defibrillator

BRIEF SUMMARY OF INDICATIONS AND IMPORTANT SAFETY INFORMATION

INDICATIONS FOR USE:

DEFIBRILLATION is a recognized means of terminating certain potentially fatal arrhythmias, such as ventricular fibrillation and symptomatic ventricular tachycardia. LIFEPAK 1000 is to be used in AED mode only on patients who are in cardiopulmonary arrest. The patient must be unresponsive, not breathing normally, and showing no signs of circulation. LIFEPAK 1000 may be used with standard defibrillation pads only on adults and children who are 8 years old or more or who weigh more than 25 kg (55 lbs). LIFEPAK 1000 may be used on children who are less than 8 years old or weigh less than 25 kg (55 lbs) with Infant/Child Reduced Energy Defibrillation Electrodes. **ECG MONITORING** is for use on conscious and unconscious patients of all ages for the purpose of ECG rhythm recognition and heart rate monitoring.

CONTRAINDICATIONS: None.

OPERATOR CONSIDERATIONS:

LIFEPAK 1000 requires operator interaction to defibrillate patient. It is intended for use by personnel who are authorized by a physician or medical director and have, at a minimum, the following skills and training: CPR training, defibrillator training equivalent to that recommended by American Heart Association, and training in the use of the LIFEPAK 1000 defibrillator. LIFEPAK 1000 is intended for use in hospital and out-of-hospital environments. Manual mode is intended for use by personnel trained in ECG recognition who want to use defibrillator to deliver a shock independent of AED mode. Operator has control over charging and delivery of shocks. ECG mode provides a nondiagnostic ECG display and is intended for use by personnel trained in ECG recognition to allow for rhythm and heart rate monitoring using standard ECG electrodes. When in ECG mode, the defibrillator's shock capability is disabled; however, LIFEPAK 1000 continues to analyze patient's ECG for potentially shockable rhythm.

GENERAL/DEFIBRILLATION WARNINGS.

SHOCK HAZARDS:

- LIFEPAK 1000 delivers up to 360 joules of electrical energy. Unless properly used, this electrical energy may cause serious injury or death. Do not attempt to operate unless thoroughly familiar with operating instructions and function of all controls, indicators, connections, and accessories.
- Clear everyone away from contact with patient, bed, and other conductive material before discharging defibrillator.
- When discharging defibrillator, do not touch electrodes.
- Do not immerse defibrillator in water or other fluids. Avoid spilling fluids on device or accessories.
- Do not disassemble defibrillator or its batteries. Contact authorized service personnel for repair.

Possible skin burns and ineffective energy delivery:

- Dried out or damaged electrodes may cause electrical arcing and patient skin burns during defibrillation. Do not use electrodes that have been removed from foil package for >24 hours or expired electrodes. Check that electrode adhesive is intact and undamaged.

Possible misinterpretation of ECG data:

- Do not analyze in a moving vehicle or move the AED during analysis. Motion artifact may affect ECG signal resulting in inappropriate shock or no shock advised message. Motion detection may delay analysis. Stop vehicle.
- Do not touch the patient or the AED during analysis.

Excessive Energy Delivery (AED mode):

- Do not use Pediatric QUIK-COMBO® electrodes; these electrodes do not attenuate the energy delivery by LIFEPAK 1000.

Implanted electrical devices:

- Defibrillation may interfere with implanted devices and cause them to malfunction. Place therapy electrodes away from implanted devices if possible.

Possible defibrillator shutdown:

- Always have access to spare, fully-charged, properly maintained battery to avoid possible device shutdown without warning.
- Replace battery when LIFEPAK 1000 displays warning of REPLACE BATTERY.

Possible device failure:

- Do not modify LIFEPAK 1000 or its batteries.

Possible explosion, fire, noxious gas or burns:

- Do not use device in presence of flammable gases or anesthetics.
- Use care when operating close to oxygen sources.
- Turn off gas source or move source away from patient during defibrillation.

Possible electrical interference or improper device performance:

- Use only parts and accessories specified by Physio-Control or Stryker. Using other manufacturers' accessories may affect performance of device or equipment in close proximity and may invalidate safety agency certification.
- Defibrillator may cause electromagnetic interference (EMI) especially during charge and energy transfers which may affect performance of equipment operating in close proximity.
- Equipment operating in close proximity may emit strong EMI or radio frequency interference (RFI) which could affect performance of device.
- Recommended distances of equipment provided in Operating Instructions.
- Safety risk and possible equipment damage.
- Keep AED away from magnetic resonance imaging (MRI) equipment as it is unsafe.

ECG MONITORING (ECG MODE) WARNINGS.

Possible delay in therapy:

- Do not attempt to connect 3-wire ECG cable to QUIK-COMBO therapy cable or any other AED.
- ECG cable is functional only with LIFEPAK 1000.

Possible misinterpretation of ECG data:

- Frequency response of screen intended only for basic ECG rhythm identification; it does not provide resolution required for pacemaker pulse visibility, accurate measurements, such as QRS duration, and ST segment interpretation. For such purposes, use ECG monitors with appropriate frequency response.

GENERAL CAUTION:

Possible equipment damage:

- Before using LIFEPAK 1000 disconnect all equipment that is not defibrillator-protected from patient.

U.S. Federal law restricts this device to sale by or on the order of a physician.

Please consult Operating Instructions at www.physio-control.com or call 800.442.1142 for complete list of indications, contraindications, warnings, cautions, potential adverse events, safety and effectiveness data, instructions for use and other important information.

For further information, please contact Stryker at 800 442 1142 (U.S.), 800 668 8323 (Canada) or visit our website at strykeremergencycare.com

Emergency Care

This document is intended solely for the use of healthcare professionals. A healthcare professional must always rely on his or her own professional clinical judgment when deciding whether to use a particular product when treating a particular patient. Stryker does not dispense medical advice and recommends that healthcare professionals be trained in the use of any particular product before using it.

The information presented is intended to demonstrate Stryker's product offerings. A healthcare professional must always refer to operating instructions for complete directions for use indications, contraindications, warnings, cautions, and potential adverse events, before using any of Stryker's products. Products may not be available in all markets because product availability is subject to the regulatory and/or medical practices in individual markets. Please contact your representative if you have questions about the availability of Stryker's products in your area. Specifications subject to change without notice.

Stryker or its affiliated entities own, use, or have applied for the following trademarks or service marks: LIFEPAK, QUIK-COMBO, Stryker. All other trademarks are trademarks of their respective owners or holders.

The absence of a product, feature, or service name, or logo from this list does not constitute a waiver of Stryker's trademark or other intellectual property rights concerning that name or logo.

GDR 3312666_B
Copyright © 2019 Stryker



Manufactured by:
Physio-Control, Inc.
11811 Willows Road NE
Redmond, WA, 98052 U.S.A.
Toll free 800 442 1142
strykeremergencycare.com

Distributed in Canada by:

Stryker Canada
2 Medicorum Place
Waterdown, Ontario
L8B 1W2
Canada
Toll free 800 668 8323